

CONFIDENTIAL

PLEASE RETURN TO R&D LAB.

011-43665

ATTN : Chief, Operations Division, Medical Staff

ENG MB-639

Chief, Engineering Staff, OC

Status of Radiological Survey and Radiation Detection Instruments  
Quarterly Report

1. During April, May and June 1959 the radiological survey and radiation detection instruments were tested to determine their operational status. A total of three instruments were inoperative and they are undergoing repair at the present time.

2. There are two Thyac Type 389 Geiger Counters inoperative because of defective high voltage power supplies. These instruments are being cannibalized and the components are being used to repair other similar instruments.

Acting

R&D/Lab/JWM/jcm (29 June 1959)

Distribution: Original and 1 - Addressee  
1 - Lab Subject  
1 - OC-E Chrono  
1 - R&D Chrono  
1 - Dev/s  
1 - Lab Pending

NOTE: The total man hours expended during the last 90 days: 6.  
This time was consumed making operational tests.

DOC	13	REV DATE	5-12-80	005632
ORIG COMP				
ORIG CLASS	C			
JUST	22	REPT DATE	2010	

PLEASE RETURN TO R&D LAB.

CONFIDENTIAL